

Comparison of response of DRI oxycodone semiquantitative immunoassay with true oxycodone values determined by liquid chromatography combined with tandem mass spectrometry: sensitivity of the DRI assay at 100 ng/ml cut-off and validity of semiquantitative value

Dixon RB, Davis B, Dasgupta A.

Journal of clinical laboratory analysis

2015; 30(3):190-195

ARTICLE IDENTIFIERS

DOI: 10.1002/jcla.21834

PMID: unavailable

PMCID: not available

JOURNAL IDENTIFIERS

LCCN: not available

pISSN: 0887-8013

eISSN: 1098-2825

OCLC ID: 13352973

CONS ID: not available

US National Library of Medicine ID: 8801384

This article was identified from a query of the SafetyLit database.